

Journal of the Association for Laboratory Automation

<http://jla.sagepub.com/>

Matrix Effects—A Challenge toward Automation of Molecular Analysis

May L. Chiu, Walson Lawi, Steven T. Snyder, Pak Kin Wong, Joseph C. Liao and Vincent Gau

Journal of Laboratory Automation 2010 15: 233

DOI: 10.1016/j.jala.2010.02.001

The online version of this article can be found at:

<http://jla.sagepub.com/content/15/3/233>

Published by:



<http://www.sagepublications.com>

On behalf of:



[Society for Laboratory Automation and Screening](#)

Additional services and information for *Journal of the Association for Laboratory Automation* can be found at:

Email Alerts: <http://jla.sagepub.com/cgi/alerts>

Subscriptions: <http://jla.sagepub.com/subscriptions>

Reprints: <http://www.sagepub.com/journalsReprints.nav>

Permissions: <http://www.sagepub.com/journalsPermissions.nav>

>> [Version of Record](#) - Jun 1, 2010

[What is This?](#)

Matrix Effects—A Challenge Toward Automation of Molecular Analysis

May L. Chiu,¹ Walson Lawi,¹ Steven T. Snyder,¹ Pak Kin Wong,² Joseph C. Liao,^{2,3}
and Vincent Gau^{1*}

¹Genefluidics, Inc., Monterey Park, CA

²The University of Arizona, Aerospace and Mechanical Engineering, Tucson, AZ

³Department of Urology, Stanford University, Stanford, CA

Keywords:

matrix effects, sample preparation, laboratory automation, microfluidics, point of care, electrokinetics, whole blood, urine, saliva

Many components in biological matrices influence the result of an analysis, affecting assay sensitivity and reproducibility. Improved matrix management becomes critical as requirements for higher assay sensitivity and increased process throughput become more demanding. There are several robotic laboratory automation systems that are commercially available, which serve to minimize matrix interference by performing purification and extraction protocols. However, there is an unmet need of inline matrix effect reduction solutions to reduce the processing time and cost for automated sample preparation. In microfluidics, effective matrix management is essential for developing fully integrated systems capable of meeting these requirements. This review surveys current biological matrix management techniques for liquid chromatography-tandem mass spectrometry (LC-MS/MS) methods and binding assays with a view toward building automatable processes. For some systems, simple sample-preparation methods, such as dilution and protein precipitation (PPT), are sufficient, whereas other systems require labor-intensive methods, such as liquid-liquid extraction (LLE) and solid-phase extraction (SPE). To achieve high throughput, PPT, LLE, and SPE have been adopted to 96-well-plate format. Online SPE has also been coupled with LC-MS/MS to automate sample preparation and analysis of urine,

plasma, and serum matrices. However, offline processing of whole blood is still required to obtain plasma and serum. The ultimate goal of implementing sample preparation to reduce matrix effects within untreated sample is to achieve reproducibility and sensitivity required by the application; therefore, inline sample preparation integrated with molecular analysis will be highly significant for laboratory automation. Electrokinetic methods have the potential of handling whole-blood, urine, and saliva samples and can be incorporated into microfluidic systems for full automation. Optimization of analysis conditions and the use of appropriate standards have likewise assisted in reducing or correcting matrix effects and will also be discussed. (JALA 2010;15:233–42)

INTRODUCTION

In bioanalysis, matrix components present in biological samples can affect the response of the analyte of interest. These phenomena, termed generally as matrix effects, can lead to inaccurate quantitation and are, therefore, important to be addressed in bioanalytical method development and validation.^{1–3} Appropriate matrix management geared toward minimizing or correcting these effects is essential in developing any reliable method. Some sample-preparation procedures have been of tremendous significance in reducing matrix effects. However, as the demand for assay sensitivity rises because of the evaluation of increasingly potent drugs and the detection of extremely low-concentration biomarkers, the need for more effective sample-preparation methods becomes more pressing. The

*Correspondence: Vincent Gau, Ph.D., Genefluidics, Inc., 2540 Corporate Place, B101, Monterey Park, CA 91754. Phone: +323.269.0900; E-mail: vgau@genefluidics.com
1535-5535/\$36.00

Copyright © 2010 by The Association for Laboratory Automation
doi:10.1016/j.jala.2010.02.001

selection of sample preparation to minimize the matrix effect is heavily determined by the specimen type and the molecular analysis platform used for analyte detection. The throughput, reagent volume, flow rate, and various protocol parameters of matrix-reduction sample preparation should be compatible with the downstream molecular analysis, and ultimately, fully integrating inline sample preparation functionality within the molecular analysis unit would be ideal for laboratory automation. Furthermore, these methods need to be fast, cost-effective, and have the ability for high throughput.⁴ It is also important to have methods that can be easily automated and integrated into an analytical module. There have been successes in coupling online solid-phase extraction (SPE) methods with liquid chromatography-tandem mass spectrometry (LC-MS/MS) methods, thus, allowing for the creation of fully automated systems for detecting analyte in urine, serum, and plasma.^{5–7} For whole-blood samples, however, removal of insolubles (blood cells or cell debris from hemolysis) is still required before sample injection.⁸ In addition, most microfluidic systems require offline sample preparation; effective sample preparation techniques that can be readily integrated into microfluidic systems and have the ability to manage whole blood samples would be particularly useful.⁹ Certain medical applications, such as point-of-care (POC) devices could stand to greatly benefit from the development of these techniques.

Matrix effects are complex and system specific. Each biological matrix presents different management challenges, and each type of analytical method is affected by matrix components differently. The focus of this review is to discuss matrix effects on LC-MS/MS methods and binding assays (immunoassays and nucleic acid hybridization assays) from urine, blood, and saliva, the most common types of biological matrices. Approaches for handling matrix effects will be discussed. Among the approaches, emphasis will be placed on current sample-preparation methods for reducing matrix effects and their status or potential for cost-effective, high-throughput processing and integration to analytical modules. Optimization of analysis conditions (e.g., LC-MS/MS parameters and assay reagents) and the use of appropriate standards have likewise assisted in reducing or correcting for matrix effects and will also be covered.

MATRIX EFFECTS OF DIFFERENT ANALYTICAL METHODS

The primary matrix effect associated with LC-MS/MS methods is ion suppression or enhancement caused by the co-eluting matrix components. This effect can be caused by both inorganic and organic endogenous substances, including salt, carbohydrates, amines, urea, lipids, peptides, and metabolites.¹⁰ Exogenous substances, such as dosing media, formulating agents, and anticoagulants, can also result in ion suppression. Ion suppression was first reported by Kobarle and Tang in the early 1990s.¹¹ Different mechanisms have been proposed to explain this phenomenon. A possible explanation is that the matrix compounds compete with

analyte for the limited charge on the droplet surfaces and, thus, affect ionization of analyte. Another possible explanation is that interfering compounds increase the droplet's viscosity and surface tension, thereby decreasing solvent evaporation rate. As a result, a lesser amount of analyte is able to reach the gas phase. It has also been suggested that nonvolatile materials in the matrix can decrease the rate of droplet formation though coprecipitation of the analyte and, therefore, prevent droplets from reaching the critical radius required for gas phase ions to be emitted.^{11–13} However, the actual mechanism is still unknown. Matrix effects are often assessed by the postcolumn infusion and the postextraction spike methods.³ The postcolumn infusion method involves constant infusion of analyte after the initial injection of an extract from the sample. It gives a qualitative assessment and can identify chromatographic regions that are likely to be influenced by matrix effects.¹⁴ In the postextraction spike method, the responses of analyte in buffered solution and a blank matrix are compared, providing a quantitative assessment of any matrix effects.¹⁵

For ligand-binding assays, nonspecific interactions of matrix components with assay components, analytes, and surfaces are the major cause of matrix effects. For instance, when a matrix component is bound nonspecifically to the antibody (an assay component) in an immunoassay, lesser antibody sites are available to bind the analyte, resulting in a weaker signal. There are also instances where matrix components (e.g., serum protein) bind nonspecifically to an analyte (e.g., hormone), decreasing antibody-antigen interaction.¹⁶ For surface plasmon resonance-based biosensors, nonspecific adsorption of matrix components has been the main type of matrix effect.¹⁷

There are very few reports on the matrix effects of nucleic acid hybridization assays. However, hybridization assays are not free of matrix effects and require sample preparation steps. The obvious matrix components that can affect a hybridization assay are the cell membranes that enclose the DNA and RNA. Cell lysis is, therefore, necessary to disrupt the cell membrane for DNA or RNA extraction. In addition, cell debris resulting from the lysis also could clot the extraction and chromatography columns and cause nonspecific binding to assay surfaces, necessitating their removal.

CHALLENGES IN MANAGING BIOLOGICAL MATRICES

Urine, blood, and saliva are among the most commonly encountered types of biological matrices. Each of these matrices has properties that present challenges in matrix management.

Urine

The management of urine matrices is difficult, because urine may contain several components that not only influence the signal response but also are of variable concentrations, dependent on diet and liquid intake. [Table 1](#) summarizes the typical constituents found in urine that can contribute toward

Table I. Constituents of infected urine that can contribute toward matrix effect

Cells	Ions (pH: 4.5–8.0)	Organic molecules	Protein	Crystal
White blood cells	Sodium	Urea, (NH ₂) ₂ CO	Immunoglobulins	Calcium oxalate, CaC ₂ O ₄
Red blood cells	Potassium	Creatinine	Tamm-Horsfall protein	Calcium phosphate, Ca ₅ (PO ₄) ₃ OH
Transitional cells	Chloride	Uric acid, C ₅ H ₄ N ₄ O ₃	Albumin	Uric acid, C ₅ H ₄ N ₄ O ₃
Squamous cells	Calcium	Citrate	Other cellular debris	Struvite, (NH ₄)MgPO ₄ ·6H ₂ O
Bacterial cells	Magnesium	Host/pathogen DNA		
Fungal cells	Ammonium	Host/pathogen RNA		
	Sulfate	Amino acids		
	Phosphate	Other organic waste		

matrix effects. Organic molecules (e.g., urea and amino acids), proteins (e.g., albumin and immunoglobulins), crystals (e.g., calcium phosphate and uric acid), and cells (white blood cells and transitional cells) are often found in urine samples and can bind nonspecifically to an assay surface (e.g., sensors and microtiter plates). Many of these compounds could cause ion suppression in LC-MS/MS analysis if co-eluted with the analyte. Furthermore, because the concentrations of these constituents are inconsistent across samples, correcting for the background signal is difficult.

Variation in composition also affects other physical properties, including pH, ionic strength, density, and viscosity. In fact, urine pH ranges from 4.5 to 8, whereas osmolality varies from 50 to 1300 mOs/kg.¹⁸ These fluctuations can influence many aspects of an analysis. In ligand-binding assays, pH and ionic strength have been shown to impact the binding affinity between the sample and assay components, resulting in response signal variations.¹⁷ Sample preparation steps could also be susceptible to differences in these conditions. For example, ion-exchanged SPE operates on the electrostatic interaction between the sample and a solid phase and is, therefore, also sensitive to pH and ionic strength. Electrokinetic phenomena, such as dielectrophoresis and electrothermal flow, which are used for sample transport and isolation in microfluidic systems, are likewise vulnerable to differences in electrical conductivity (ion concentration)

and pH. Aside from pH and ionic strength, viscosity could be influenced by cell concentration within a sample, thereby affecting the operation of the microfluidic system performing the analysis.

Blood

Whole blood has one of the most complex matrices (Fig. 1). Many matrix components can affect the signal response of a bioanalytical process. Serum proteins often bind nonspecifically to analytes or a surface resulting in reduced sensitivity.¹⁷ Endogenous phospholipids have been identified as a major source of matrix effects in LC-MS type of analysis.¹⁹ Anticoagulant, which is used in blood collection to prevent coagulation of blood, has been found to have significant effect on extraction and LC-MS/MS analysis.²⁰ The high viscosity of blood also alters the binding efficiency and specificity for detection. Some informative biomarkers, such as transplant rejection markers, are embedded inside distinct subsets of lymphocytes at extremely low concentrations. Matrix management for detecting this type of analyte is even more challenging, because even low levels of residue matrix components can skew the analysis.

A final complicating factor in blood matrix management is that removal of blood cells or cell debris from hemolysis is necessary for analysis. Centrifugation and filtration, a pair

Plasma	Cells					
	Erythrocytes	Leukocytes				Platelets
Water	~8 μm	7-20 μm				2-5 μm
Proteins		Granulocytes		Agranulocytes		
Electrolytes						
Glucose	Rouleaux	Basophils	Eosinophils	Neutrophils	Lymphocytes	Monocytes
Amino acids						
Lipids	Hemolyzed					
Hormones	Crenated					
Waste						
Gases						

Figure 1. Common contributors to the matrix effects in blood.

of conventional techniques for blood cell separation, are difficult to implement in an automated microfluidic format.

Saliva

Saliva is an attractive fluid for biological analysis, because it allows for noninvasive sampling procedures. Saliva assays have been used since early 1980s primarily for the detection of small molecule drugs and steroids.²¹ Despite its attractively easy sampling procedure, saliva has been less commonly used as a sample source, and results from saliva analyses tend to be viewed as less reliable than analyses performed using other biological fluids. This is mainly because saliva is a difficult matrix to manage. Saliva matrix has significant effects on sampling. It has been shown that saliva pH is influenced by flow rate, which affects analyte excretion. For example, Kato et al. showed that stimulated collections of saliva increased the saliva pH and resulted in lower concentrations of cocaine.²² Jenkins et al. had similar findings in their work on comparing the concentrations of heroin and cocaine in saliva with concentrations of the drugs in plasma.²³ Because there is always lesser analyte in saliva than in urine or blood, stimulation is often necessary to obtain a sufficient volume of sample for analysis. Careful method validation for saliva sampling is therefore essential for using saliva from the sample specimen. In addition to complications arising because of pH, results obtained from saliva can also be affected by drug residues in oral or nasal cavities.²⁴

Immunoassays of saliva are also subject to matrix effects. Fulton et al. reported that mucins and alpha amylase in saliva matrix could cause suppression of antibody binding in some radioimmunoassays for steroid detection.²⁵ The results shown by Mitchell and Lowe also demonstrated that saliva matrix caused suppression of antibody-binding enzyme in immunoassays for testosterone,²⁶ and removal of saliva medium was necessary to eliminate the matrix effects.

CURRENT APPROACHES IN DEALING WITH MATRIX EFFECTS

Sample Preparation

Appropriate sample preparation can help eliminate or minimize matrix effects. Several technical approaches, including optical,^{27,28} magnetic,²⁹ mechanical,³⁰ and electrical³¹ approaches, have been demonstrated to be effective for sample manipulation and matrix effect reduction. Matrix effects are system specific. Depending on the matrix type, detection method, characteristics of the analyte, and the limit of detection requirement, different sample-preparation methods are required. For some systems, simple dilution is sufficient for accurate analysis of the analyte.³²⁻³⁴ Perhaps, dilution, the most basic of preparatory methods, is effective in situations where matrix effects decrease at a rate higher than that of the analyte response as the sample is diluted. These are the systems that tend to have a less demanding requirement on detection limit and are often characterized by samples containing abundant analyte. For example, many small-molecule drug analyses

involve high analyte concentrations in urine, which allow for a simple dilution of the sample before analysis. There are also systems with relatively low detection limits in the pico- or femto-molar range, where dilution was sufficient because of particularly robust detection methods.^{33,35,36} For most systems, however, dilution is not suitable, because analyte concentrations in a diluted sample are too low for detection. Table 2 is a summary of the most commonly used sample-preparation methods to reduce the matrix effects of urine, whole blood, and saliva.

Isolation and extraction are often used to remove or minimize matrix components and, thereby, eliminate or minimize matrix effects. For samples with insolubles (e.g., cells), such separation of a soluble material from an insoluble material is usually the first step in sample preparation.³⁷ Centrifugation is often used to achieve separation. For example, cell lysates can be centrifuged to remove cell debris, whereas DNA, RNA, and proteins can subsequently be collected from the supernatant for analysis or further purification. Centrifugation is frequently used for processing blood samples, because the removal of blood cells from whole blood is often necessary for analysis or further cleanup. Centrifugation is also used as an element in other sample-cleanup methods to achieve more efficient separation for small-volume samples. For example, various types of SPEs (e.g., size exclusion, reverse phase, ion exchange, and immunoprecipitation) are available in spin (centrifuge) column format. Filtration with the context of centrifugation has also been a popular tool for cleaning proteins.³⁸ Although widely used, centrifugation is labor intensive and bulky. Implementing centrifugation in an automated microfluidic format or finding practical equivalent alternatives are ongoing challenges.

Precipitation is another straightforward technique and functions by creating insolubles. By subjecting the sample to a solvent with low solubility for select components, the insoluble components within the solvent precipitate and can be subsequently separated from the soluble phase, often by centrifugation. Protein precipitation (PPT) is a popular sample-preparation method for MS-based analysis³⁹⁻⁴² because of its simplicity. PPT in 96-well filter plate format also allows for high-throughput and automated sample preparation.⁴³ However, PPT does not always sufficiently remove salt and

Table 2. Summary of commonly used specimen-specific sample-preparation methods

Specimen type	Matrix effect reduction sample preparation
Urine	SPE ^{103,104} Centrifugation ¹⁰⁵ Dilution, protein precipitation ¹⁰⁶
Blood	Extraction ¹⁰⁷ Surface coating ⁹⁸
Saliva	pH adjustment, surface coating ¹⁰⁸ Chemical treatment ¹⁰⁹ Magnetic bead ¹¹⁰

endogenous compounds, such as lipids, phospholipids, and fatty acids.¹⁹, which could cause ion suppression in LC-MS/MS analysis. Choice of solvent could improve extract purity. For instance, Chambers et al. demonstrated that acetonitrile is more efficient in removing phospholipids than methanol.¹⁹ PPT has also been used in combination with other strategies to achieve the desired sample purity.⁴⁴

Liquid-liquid extraction (LLE) is also one of the most commonly used sample-preparation techniques. This method separates compounds based on differing solubilities in two immiscible or partially soluble contacting liquid phases, usually an aqueous and a nonpolar organic phase. LLE is widely used for cleanup of samples containing lower-molecular-weight analytes, such as drugs and hormones.^{45–50} Although infrequent, LLE is sometimes used for extraction of proteins when two aqueous phases are used.⁵¹ LLE usually yields rather clean extracts and gives high reproducibility. However, the process often necessitates preconcentration before analysis, requires large amounts of costly toxic solvents, and is time consuming. LLE is also difficult to automate.⁵² There has been some progress in implementing LLE methods in the 96-well-plate format, but automating efficiently the mixing and phase separation steps remains a challenge.^{4,53,54}

SPE is another widely used sample-extraction method. It is faster and requires lesser solvent compared with LLE. SPE involves a partitioning of compounds between mobile (liquid) and stationary (solid) phases.⁵⁵ The analyte is first retained by the stationary phase through hydrophobic (reverse-phase SPE) or electrostatic (ion-exchanged SPE) interaction, whereas the matrix components are washed off. After the removal of the sample matrix, the retained analyte on the stationary phase is eluted with a solvent possessing a greater affinity for the analyte. It is widely used for extraction of both low-molecular-weight compounds and proteins and is effective in eliminating matrix effects for a variety of systems.^{56–60} In fact, Chambers et al.¹⁹ conducted a study comparing cleanliness and matrix effects of samples extracted from plasma using PPT, LLE, and various types of SPE methods. The results demonstrated that mixed-mode SPE (both reverse phase and ion exchange) provided the cleanest extract, leading to considerable reduction in matrix effects.¹⁹ Dams et al.³² also found SPE to be a more extensive cleanup method compared with PPT in a study on LC-MS/MS analysis of illicit drugs. However, ion suppression was still observed in the sample extract when SPE was used; this was because of the preconcentration step in SPE, which raises the concentration of the nonremoved interfering substance along with the analyte.³² Much progress has been made in automating SPE. SPE in 96-well-plate format has been commonly used.⁶¹ Online SPE, often coupled with LC-MS/MS, is increasingly getting popular.^{5–7} It is generally considered to be faster and more cost-effective compared with the 96-well-plate format.⁴ Different extraction support materials have been developed to allow for direct injection of urine, plasma, serum, or extracts. However, for blood samples, isolation of blood cells is still required before performing SPE, because blood cells will inevitably clot the extraction column.

In the past two decades, there has been a growing interest in extracting analyte through antibody and antigen interaction.⁶² Immunoaffinity SPE consists of a stationary phase of immobilized antibodies specific for the target analyte. There are a variety of stationary phase materials, including centrifugal filters,⁶³ magnetic beads,⁶⁴ polymer beads,⁶⁵ and silica.⁶⁶ The analyte is retained through its specific interaction with immobilized antibodies after other compounds are removed. Subsequently, a low-pH buffer, usually between pH 2 and 4, is used to disrupt the antibody-antigen interaction, thereby eluting the analyte. This approach has been successfully used in extracting drugs from plasma and urine,^{67–71} though in many cases, additional sample-preparation steps, such as PPT, were necessary before the immunoaffinity extraction step. The advantage of this approach is the ability to separate structurally similar analytes. However, the process is costly. This is especially true for high-throughput processes, because antibodies are expensive, require refrigeration, and degrade relatively quickly.

A powerful approach for matrix management is alternating current (AC) electrokinetics, which is the motion of bulk fluids or embedded particles induced by nonuniform AC electric fields. AC electrokinetics, which is especially effective in the micro and nano domains, is ideal for manipulating biological targets, such as nucleic acids, proteins, and cells, because of the proper length scale matching.^{72–74} At a small scale, only a small voltage is required to generate a large electric field. Other advantages of AC electrokinetics include label-free manipulation (in contrast to magnetic particle-based manipulation), well-established techniques for fabricating microelectrodes (easy integration with other components), and low voltage requirement (which avoids electrolysis and facilitates portable applications). Examples of AC electrokinetic phenomena include dielectrophoresis (DEP), AC electrothermal flow (ACET), and AC electroosmosis (ACEO).⁷⁵ DEP is a result of the net force experienced by a dipole under nonuniform electric field. It can be applied for trapping biological objects, such as bacteria, DNA, and viral particles^{70,71,76–79} (Fig. 2A). DEP is especially effective in short range because of the rapid decay of the electric field gradient. ACET arises from the temperature gradient in the fluid medium generated by Joule heating. The temperature gradient induces local changes in the conductivity and permittivity of the fluid medium. These gradients further interact with the electric field and generate bulk fluid forces (Fig. 2B). For example, a conductivity gradient produces Coulomb forces, whereas a permittivity gradient produces dielectric forces. This results in long-term fluid motion. ACEO is an electrokinetic phenomenon observed at frequency ranges below 1 MHz.^{80–82} When a potential is applied to the electrode, the electric field causes charges to accumulate on the electrode surface, which changes the charge density near the electrode surface and forms an electric double layer (EDL). The EDL interacts with the tangential component of the electric field to induce bulk fluid motion. In an alternating dielectric field, both sign of the charges in the EDL and




Figure 2. AC electrokinetic phenomena for matrix management. (A) In dielectrophoresis, a target cell or molecule is polarized to induce a dipole. If the electric field is nonuniform, the dipole experiences a net force, which results in translational motion of the target. (B) In AC electrothermal flow, Joule heating induces gradients in temperature and electrical properties of the fluid, which results in bulk fluid motion. (C) In AC electroosmosis, the induced electric double layer on the electrode surface interacts with the electric field. The resulting force drives the motion of the fluid near the electrode surface.

direction of the tangential component of the electric field change (Fig. 2C). Therefore, the direction of the resultant force on the fluid remains the same and generates a net fluid movement. The bulk fluid flow allows long-range manipulations of bioparticles with low applied voltage (< 10 V). Each electrokinetic phenomenon has different effective range and optimal operating conditions. Recently, hybrid electrokinetics, the combination of long-range electrokinetic induced fluid motion (e.g., ACEO and ACET) and short-range electrokinetic trapping force (e.g., DEP), has been developed to take advantage of the effective regimes of different electrokinetic phenomena.^{83–85}

Optimization of Analytical Method Parameters and Use of Sample Standards

In addition to proper sample preparation, a large amount of effort has been expended to optimize analytical methods and use appropriate standards toward eliminating or reducing matrix effects.

Liquid Chromatography-Tandem Mass Spectrometry. Atmospheric pressure chemical ionization (APCI) and electrospray ionization (ESI) are the most popular atmospheric pressure ionization techniques. The ionizing mechanisms are different, and their susceptibility to matrix effects also varies. Several studies have found APCI to be less susceptible to matrix effects.^{86,87} For instance, Souverain et al. found APCI to be less susceptible to a matrix effect for methodone extracted using different sample-cleaning methods.⁸⁶ However, some matrix effects persist while using APCI. Furthermore,

the design of the source interface also influences its susceptibility to matrix effects.⁸⁸

The ESI flow rate also has been found to influence the degree of ion suppression. Studies demonstrate a dramatic reduction of matrix effects with a decrease in ESI flow rate to nanoliter per minute.^{89,90} Some researchers have suggested that this occurs because of the decrease in the initial charged droplet size resulting from the lower flow rate. Consequently, fewer uneven fission events and lesser solvent evaporation are required for ion release in the gas phase.⁹¹ Because uneven fission process enhances the surface-activated matrix components to compete with analytes for a limited number of surface charges, less uneven fission will minimize competition and lead to a stronger analyte signal.⁹²

A straightforward means for correcting matrix effects is to use an appropriate internal standard. For the internal standard to be meaningful, the internal standard and the analyte need to co-elute to ensure that they are ionized under the same conditions. Therefore, stable isotope-labeled (SIL) internal standard is the commonly used internal standard in LC-MS/MS detection methods. It has even been a common belief that the use of an isotopically labeled internal standard corrects for almost all matrix effects. However, this has been shown to be untrue. Wang et al. demonstrated that a difference in retention time between the analyte and the SIL internal standard, caused by the deuterium isotope effect, results in a differing degree of ion suppression between the sample and the internal standard.⁹³ Furthermore, the analyte and co-eluting SIL internal standard could also have mutual effects (suppression or enhancement) on each other's response.^{94,95} Therefore, careful assessment should be made during method development. In addition, SIL internal standard is also costly for multicomponent analysis, because an SIL internal standard is required for each analyte.

Improved chromatography efficiency can also help to separate matrix components from the analyte more effectively. Matuszewski et al. demonstrated that, by providing longer chromatographic retention of the analyte, the plasma matrix components were well separated from the analyte, thereby eliminating ion suppression.⁹⁶ A prolonged high-performance liquid chromatography gradient separation was used to resolve interfering peaks, allowing for more reliable integration of both the internal standards and the analytes of interest.⁸⁸ Hydrophilic interaction chromatography has also been found to be more appropriate for polar analyte analysis.⁹⁷

Ligand-Binding Assays. For ligand-binding assays, buffer composition, pH, and ionic strength are critical in reducing nonspecific interactions. For example, Situ et al.⁹⁸ were able to reduce nonspecific interactions between the serum component and the surface up to 94% by optimizing buffer pH and composition. Johansson and Hellenas¹⁷ also demonstrated that appropriate buffer pH and ionic strength conditions can help reduce matrix effects caused by nonspecific adsorption of urine matrix components.

Displacing agents are sometimes used to minimize matrix effects of nonspecific binding between serum protein and analyte. In a progesterone assay, Ratcliffe et al. used unrelated steroids to compete for serum protein-binding sites, thereby displacing and freeing the progesterone for detection.⁹⁹ The drawback of using displacing agents is the possibility of cross-reaction.¹⁰⁰

Adequate washing is important in removing nonspecifically adsorbed components on a surface. Insufficient washing leads to high background. Therefore, homogeneous assays, which lack washing and separation steps, are often more susceptible to matrix effects.¹⁰¹

In addition, appropriate choice of assay components is also essential in reducing matrix effects. For example, Meyer et al. demonstrated that, by using a highly specific antibody against urodilatin, matrix effects were not observed in the RIA for urinary and plasma urodilatin detection even without extraction.³³ It is also important to be mindful that the choice of assay components need to be tailored for different systems based on factors, such as analyte, matrix type, and detection method. For instance, lysozyme was a great reporter enzyme for urine samples in enzyme-multiplied immunoassay technique assays but ineffective for serum samples, because bacterial cells (substrates for lysozyme) agglutinate the serum.¹⁰²

To compensate for matrix effects, calibration curves spiked in sample matrix or prepared in phosphate buffer with addition of urea and creatinine have been used. However, lot-to-lot variation could still lead to inaccurate results, and need to be evaluated during method development and validation.

CONCLUSION

Appropriate matrix management is essential for the development of a sensitive and reliable bioanalytical method. This review focused on the management of urine, blood, and saliva for LC-MS/MS and binding assays. Reliable systems have been created using conventional sample-preparation methods, including dilution, centrifugation, precipitation, LLE, and SPE. However, these methods are either not sufficient for highly sensitive assays or are labor intensive and time consuming. Many are difficult to automate and integrate with other steps and hinder full automation. Online SPE has been coupled with LC-MS/MS to achieve full automation for urine, plasma, and serum matrices; however, processing of whole blood is still required to extract plasma and serum. Electrokinetic methods, which can be readily incorporated into microfluidic systems, have the capability to manipulate particles of different sizes and shapes, including cells, proteins, and DNA molecules. More effort should be expended on developing electrokinetic methods for matrix management.

ACKNOWLEDGMENTS

This work was supported by National Institute of Allergy and Infectious Diseases (grants U01 AI075565-01 and U01 AI082457) and National Institute of Dental and Craniofacial Research (grant U01 DE17790).

Competing interests statement: The authors certify that all financial and material support for this research and work are clearly identified in the manuscript.

REFERENCES

1. Viswanathan, C. T.; Bansal, S.; Booth, B.; DeStefano, A. J.; Rose, M. J.; Sailstad, J.; Shah, V. P.; Skelly, S. P.; Swann, P. G.; Weiner, W. Workshop/conference report—Quantitative bioanalytical methods validation and implementation: best practices for chromatographic and ligand binding assays. *AAPS J.* **2007**, *9*, E30–E42.
2. “Guidance for Industry Bioanalytical Method Validation.” U.S. Department of Health and Human Services Food and Drug Administration, Center for Drug Evaluation and Research (CDER), May 2001.
3. Eeckhaut, A. V.; Lanckmans, K.; Sarre, S.; Smolders, I.; Michotte, Y. Validation of bioanalytical LC–MS/MS assays: evaluation of matrix effects. *J. Chromatogr. B* **2009**, *877*, 2198–2207.
4. Xu, R. N.; Fan, L.; Rieser, M. J.; El-Shourbagy, T. A. Recent advances in high-throughput quantitative bioanalysis by LC–MS/MS. *J. Pharm. Biomed. Ther.* **2007**, *44*, 342–355.
5. Barrett, Y. C.; Akinsanya, B.; Chang, S.-Y.; Vesterqvist, O. Automated on-line SPE LC–MS/MS method to quantitate 6beta-hydroxycortisol and cortisol in human urine: use of the 6beta-hydroxycortisol to cortisol ratio as an indicator of CYP3A4 activity. *J. Chromatogr. B* **2005**, *821*, 159–165.
6. Zhou, S.; Zhou, H.; Larson, M.; Miller, D. L.; Mao, D.; Jiang, X.; Naidong, W. High-throughput biological sample analysis using on-line turbulent flow extraction combined with monolithic column liquid chromatography/tandem mass spectrometry. *Rapid Commun. Mass Spectrom.* **2005**, *19*, 2144–2150.
7. Niederländer, H. A. G.; Koster, E. H. M.; Hilhorst, M. J.; Metting, H. J.; Eilders, M.; Ooms, B.; de Jong, G. J. High throughput therapeutic drug monitoring of clozapine and metabolites in serum by on-line coupling of solid phase extraction with liquid chromatography–mass spectrometry. *J. Chromatogr. B* **2006**, *834*, 98–107.
8. Ceglarek, U.; Lembecke, J.; Fiedler, G. M.; Helmut, M. W.; Witzigmann, H.; Hauss, J. P.; Thiery, J. Rapid simultaneous quantification of immunosuppressants in transplant patients by turbulent flow chromatography combined with tandem mass spectrometry. *Clin. Chim. Acta* **2004**, *346*, 181–190.
9. Kim, J.; Johnson, M.; Hill, P.; Gale, B. K. Microfluidic sample preparation: cell lysis and nucleic acid purification. *Integr. Biol.* **2009**, *1*, 574–586.
10. Antignac, J.-P.; de Wasch, K.; Monteau, F.; de Brabander, H.; Andre, F.; le Bizec, B. The ion suppression phenomenon in liquid chromatography–mass spectrometry and its consequences in the field of residue analysis. *Anal. Chim. Acta* **2005**, *29*, 129–136.
11. Kebarle, P.; Tang, L. From ions in solution to ions in the gas phase—the mechanism of electrospray mass spectrometry. *Anal. Chem.* **1993**, *65*, 972A–986A.
12. King, R.; Bonfiglio, R.; Fernandez-Metzler, C.; Miller-Stein, C.; Olah, T. Mechanistic investigation of ionization suppression in electrospray ionization. *J. Am. Soc. Mass Spectrom.* **2000**, *11*, 942–950.
13. Reilly, P. T. A.; Lazar, A. C.; Gieray, R. A.; Whitten, W. B.; Ramsey, J. M. The elucidation of charge-transfer-induced matrix effects in environmental aerosols via real-time aerosol mass spectral analysis of individual airborne particles. *Aerosol Sci. Tech.* **2000**, *33*, 135–152.

14. Bonfiglio, R.; King, R. C.; Olah, T. V.; Merkle, K. The effects of sample preparation methods on the variability of the electrospray ionization response for model drug compounds. *Rapid Commun. Mass Spectrom.* **1999**, *13*, 1175–1185.
15. Matuszewski, B. K.; Constanzer, M. L.; Chavez-Eng, C. M. Strategies for the assessment of matrix effect in quantitative bioanalytical methods based on HPLC-MS/MS. *Anal. Chem.* **2003**, *75*, 3019–3030.
16. Schioler, V.; Thode, J. Six direct radioimmunoassays of estradiol evaluated. *Clin. Chem.* **1988**, *34*, 949–952.
17. Johansson, M. A.; Hellenas, K. E. Matrix effects in immunobiosensor determination of clenbuterol in urine and serum. *Analyst* **2004**, *129*, 438–442.
18. Sviridov, D.; Hortin, G. L. Urine albumin measurement: effects of urine matrix constituents. *Clin. Chim. Acta* **2009**, *404*, 140–143.
19. Chambers, E.; Wagrowski-Diehl, D. M.; Lu, Z.; Mazzeo, J. R. Systematic and comprehensive strategy for reducing matrix effects in LC/MS/MS analyses. *J. Chromatogr. B* **2007**, *852*, 22–34.
20. Chin, C.; Zhang, Z. P.; Karnes, H. T. A study of matrix effects on an LC/MS/MS assay for olanzapine and desmethyl olanzapine. *J. Pharmaceut. Biomed* **2004**, *35*, 1149–1167.
21. Lac, G. Saliva assays in clinical and research biology. *Pathol. Biol.* **2001**, *49*, 660–667.
22. Kato, K.; Hills Grove, M.; Weinhold, L.; Gorelick, D. A.; Darwin, W. D.; Cone, E. J. Cocaine and metabolite excretion in saliva under stimulated and nonstimulated conditions. *J. Anal. Toxicol.* **1993**, *17*, 338–341.
23. Jenkins, A. J.; Oyler, J. M.; Cone, E. J. Comparison of heroin and cocaine concentrations in saliva with concentrations in blood and plasma. *J. Anal. Toxicol.* **1995**, *19*, 359–374.
24. O'Neal, C. L.; Crouch, D. J.; Rollins, D. E.; Fatah, A.; Cheever, M. L. Correlation of saliva codeine concentrations with plasma concentrations after oral codeine administration. *J. Anal. Toxicol.* **1999**, *23*, 452–459.
25. Fulton, A.; Chan, S.; Coleman, G. Effect of salivary proteins on binding curves of three radioimmunoassay kits: Amerlex-M Progesterone, Amerlex Cortisol, and Biodata Testosterone. *Clin. Chem.* **1989**, *35*, 641–644.
26. Mitchell, J. S.; Lowe, T. E. Matrix effects on an antigen immobilized format for competitive enzyme immunoassay of salivary testosterone. *J. Immunol. Methods* **2009**, *349*, 61–66.
27. Eriksson, E.; Scrimgeour, J.; Granéli, A.; Ramser, K.; Wellander, R.; Enger, J.; Hanstorp, D.; Goksör, M. Optical manipulation and microfluidics for studies of single cell dynamics. *J. Opt. A Pure Appl. Opt.* **2007**, *9*, S113–S121.
28. Oroszi, László; Dér, András; Kirei, H. Manipulation of microfluidic flow pattern by optically controlled electroosmosis. *Microfluid. Nanofluid.* **2009**, *6*, 565–569.
29. Bu, M.; Christensen, T. B.; Smistrup, K.; Wolff, A.; Hansen, M. F. Characterization of a microfluidic magnetic bead separator for high-throughput applications. *Sensors Actuators A Physical* **2008**, *145–146*, 430–436.
30. Yi, C.; Li, C.-W.; Ji, S.; M. Yang. Microfluidics technology for manipulation and analysis of biological cells. *Anal. Chim. Acta* **2006**, *560(1–2, 23)*, 1–23.
31. Lin, R.; Burke, D. T.; Burns, M. A. Addressable electric fields for size-fractionated sample extraction in microfluidic devices. *Anal. Chem.* **2005**, *77(14)*, 4338–4347.
32. Dams, R.; Huestis, M. A.; Lambert, W. E.; Murphy, C. M. Matrix effect in bio-analysis of illicit drugs with LC-MS/MS: influence of ionization type, sample preparation, and biofluid. *J. Am. Soc. Mass Spectrom.* **2003**, *14*, 1290–1294.
33. Meyer, M.; Armbruster, F. P.; Maier, I.; Schwing, J.; Haller, C.; Missbichler, A.; Knut, A.; Forssmann, W.-G.; Woloszczuk, W. Urinary and plasma urodilatin measured by a direct RIA using a highly specific antiserum. *Clin. Chem.* **1998**, *44*, 2524–2529.
34. Cooke, D. G.; Binnie, J. E.; Blackwell, L. F. Validation of a reference ELISA for estrone glucuronide using urine samples normalized by dilution to a constant rate of urine production. *Steroids* **2007**, *72*, 580–591.
35. Fang, W.; Wang, J.; Liao, W.; Zimmermann, B. G.; Wong, D.; Ho, C.-H. Electrochemical detection of low-copy number salivary RNA based on specific signal amplification with a hairpin probe. *Nucleic Acids Res.* **2008**, *36*, e65.
36. Gau, V.; Wong, D. Oral fluid nanosensor test (OFNASET) with advanced electrochemical-based molecular analysis platform. *Ann. N.Y. Acad. Sci.* **2007**, *1098*, 401–410.
37. Belter, P. A.; Cussler, E. L.; Hu, W.-S. *Bioseparations: Downstream Processing for Biotechnology*. John Wiley & Sons: New York, NY; 1988.
38. Yuan, M.; Carmichael, W. W.; Hilborn, E. D. Microcystin analysis in human sera and liver from human fatalities in Caruaru, Brazil 1996. *Toxicol.* **2006**, *48*, 627–640.
39. Yang, Y.; Li, H.; Gao, K.; Liu, M.; Sun, Y.; Yan, T.; Fawcett, J. P.; Cui, Y.; Gu, J. Simultaneous quantitation of dexamethasone palmitate and dexamethasone in human plasma by liquid chromatography/tandem mass spectrometry. *J. Chromatogr. B* **2007**, *862*, 119–124.
40. Le Saux, T.; Chhun, S.; Rey, E.; Launay, O.; Weiss, L.; Viard, J.-P.; Pons, G.; Jullien, V. Quantification of seven nucleoside/nucleotide reverse transcriptase inhibitors in human plasma by high-performance liquid chromatography with tandem mass-spectrometry. *J. Chromatogr. B* **2008**, *865*, 81–90.
41. Cao, X.; Yao, Z.; Chen, H.; Dai, Y.; Sun, P.; Ye, W.; Yao, X. Development and validation of a liquid chromatography tandem mass spectrometry assay for the quantification of methyl protodioscin in rat plasma: application to a pharmacokinetic study. *Biomed. Chromatogr.* **2007**, *22*, 408–413.
42. Flarakos, J.; Luo, W.; Aman, M.; Svinarov, S.; Gerber, N.; Vouros, P. Quantification of risperidone and 9-hydroxyrisperidone in plasma and saliva from adult and pediatric patients by liquid chromatography-mass spectrometry. *J. Chromatogr. A* **2004**, *1026*, 175–183.
43. Walter, R. E.; Cramer, J. A.; Tse, F. L. S. Comparison of manual protein precipitation (PPT) versus a new small volume PPT 96-well filter plate to decrease sample preparation time. *J. Pharmaceut. Biomed.* **2001**, *25*, 331–337.
44. Xue, Y. J.; Akinsanya, J. B.; Liu, J.; Unger, S. E. A simplified protein precipitation/mixed-mode cation-exchange solid-phase extraction, followed by high-speed liquid chromatography/mass spectrometry, for the determination of a basic drug in human plasma. *Rapid Commun. Mass Spectrom.* **2006**, *20*, 2660–2668.
45. Ji, H. Y.; Park, E. J.; Lee, K. C.; Lee, H. S. Quantification of doxazosin in human plasma using hydrophilic interaction liquid chromatography with tandem mass spectrometry. *J. Sep. Sci.* **2008**, *31*, 1628–1633.
46. Buhrman, D. L.; Price, P. I.; Rudewicz, P. J. Quantitation of SR 27417 in human plasma using electrospray liquid chromatography-tandem

- mass spectrometry: a study of ion suppression. *J. Am. Soc. Mass. Spectrom.* **1996**, *7*, 1099–1105.
47. Lee, H.-S.; Kang, I.-M.; Lee, H.-W.; Seo, J.-H.; Ryu, J.-H.; Choi, S.-J.; Lee, M.-J.; Jeong, S. Y.; Cho, Y.-W.; Lee, K.-T. Development and validation of high performance liquid chromatography-tandem mass spectrometry for the determination of etodolac in human plasma. *J. Chromatogr. B* **2008**, *863*, 158–162.
 48. Nirogi, R. V. S.; Kandikere, V. N.; Shukla, M.; Mudigonda, K.; Maurya, S.; Boosi, R.; Anjaneyulu, Y. Development and validation of a sensitive liquid chromatography/electrospray tandem mass spectrometry assay for the quantification of olanzapine in human plasma. *J. Pharmaceut. Biomed* **2006**, *41*, 935–942.
 49. Chen, K.-B.; Cai, M.-Q.; Chen, X.-H.; Jin, M.-C. Quantitative analysis of triptolide in human whole blood by LC-APCI-IT-MS-MS. *Chromatographia* **2008**, *67*, 225–230.
 50. Dighe, A. S.; Sluss, P. M. Improved detection of serum estradiol after sample extraction procedure. *Clin. Chem.* **2003**, *50*, 764–766.
 51. Sarmento, M. J.; Pires, M. J.; Cabral, J. M. S.; Aires-Barros, M. R. Liquid-liquid extraction of a recombinant protein, cytochrome b5, from an impure extract using aqueous two-phase systems. *Bioproc. Biosyst. Eng* **1997**, *16*, 295–297.
 52. Pawlizyn, J. In *Sampling and Sample Preparation for Field and Laboratory*, Vol. 37; Elsevier: Amsterdam, The Netherlands; 2002; pp 727.
 53. Wang, P. G.; Wei, J. S.; Kim, G.; Chang, M.; El-Shourbagy, T. Validation and application of a high-performance liquid chromatography-tandem mass spectrometric method for simultaneous quantification of lopinavir and ritonavir in human plasma using semi-automated 96-well liquid-liquid extraction. *J. Chromatogr. A* **2006**, *1130*, 302–307.
 54. Xu, X. S.; Yan, K. X.; Song, H.; Lo, M. W. Quantitative determination of a novel dual PPAR α/γ agonist using on-line turbulent flow extraction with liquid chromatography-tandem mass spectrometry. *J. Chromatogr. B* **2005**, *814*, 29–36.
 55. Berrueta, L. A.; Gallo, B.; Vicente, F. A review of solid phase extraction: basic principles and new developments. *Chromatographia* **1995**, *40*, 474–483.
 56. Matsuura, K.; Ohmori, T.; Nakamura, M.; Itoh, Y.; Hirano, K. A simple and rapid determination of valproic acid in human plasma using a non-porous silica column and liquid chromatography with tandem mass spectrometric detection. *Biomed. Chromatogr.* **2008**, *22*, 387–393.
 57. Tranfo, G.; Paci, E.; Sisto, R.; Pignini, D. Validation of an HPLC/MS/MS method with isotopic dilution for quantitative determination of trans, trans-muconic acid in urine samples of workers exposed to low benzene concentrations. *J. Chromatogr. B* **2008**, *867*, 26–31.
 58. Mistri, H. N.; Jangid, A. G.; Pudage, A.; Shrivastav, P. HPLC-ESI-MS/MS validated method for simultaneous quantification of zopiclone and its metabolites, N-desmethyl zopiclone and zopiclone-N-oxide in human plasma. *J. Chromatogr. B* **2008**, *862*, 137–148.
 59. Lindegardh, N.; Annerberg, A.; White, N. J.; Day, N. P. J. Development and validation of a liquid chromatographic-tandem mass spectrometric method for determination of piperazine in plasma stable isotope labeled internal standard does not always compensate for matrix effects. *J. Chromatogr. B* **2007**, *862*, 227–236.
 60. Naik, G. O.; Moe, G. W. Use of a solid-phase extraction with radioimmunoassay to identify the proportional bias of clinical B-type natriuretic peptide immunoassay: the impact of plasma matrix and antibody multispecificity. *Clin. Chem. Lab. Med.* **2007**, *45*, 1353–1359.
 61. Jiang, H.; Li, Y.; Pelzer, M.; Cannon, M. J.; Randlett, C.; Junga, H.; Jiang, X.; Ji, Q. J. Determination of molindone enantiomers in human plasma by high-performance liquid chromatography-tandem mass spectrometry using macrocyclic antibiotic chiral stationary phases. *J. Chromatogr. A* **2008**, *119*, 230–238.
 62. Stevenson, D. Immuno-affinity solid-phase extraction. *J. Chromatogr. B* **2000**, *745*, 39–48.
 63. Haasnoot, W.; Kemmers-Voncken, A. Immunofiltration as sample cleanup for the immunochemical detection of β -agonist in urine. *Analyst* **2002**, *127*, 87–92.
 64. Hansel, T. T.; De Vries, I. J.; Iff, T.; Rihs, S.; Wandzilak, M.; Betz, S.; Blaser, K.; Walker, C. An improved immunomagnetic procedure for the isolation of highly purified human blood eosinophils. *J. Immunol. Methods* **1991**, *145*, 105–110.
 65. Hayashi, Y.; Shono, F.; Yamamoto, S.; Takasaki, W.; Nakagawa, A.; Watanabe, K.; Yamashita, K.; Miyazaki, H. Immunoaffinity purification of 11-dehydro-thromboxane B2 from human urine and plasma for quantitative analysis by radioimmunoassay. *Anal. Biochem.* **1990**, *187*, 151–159.
 66. Rashid, B. A.; Kwasowski, P.; Stevenson, D. Solid phase extraction of clenbuterol from plasma using immunoaffinity followed by HPLC. *J. Pharm. Biomed.* **1999**, *21*, 635–639.
 67. Degan, P.; Shigenaga, M. K.; Park, E.-M.; Alperin, P. E.; Ames, B. N. Immunoaffinity isolation of urinary 8-hydroxy-2'-deoxyguanosine and 8-hydroxyguanine and quantitation of 8-hydroxy-2'-deoxyguanosine in DNA by polyclonal antibodies. *Carcinogenesis* **1991**, *12*, 865–871.
 68. Rashid, B. A.; Aherne, G. W.; Katmeh, M. F.; Kwasowski, P.; Stevenson, D. Determination of morphine in urine by solid-phase immunoextraction and high-performance liquid chromatography with electrochemical detection. *J. Chromatogr. A* **1998**, *797*, 245–250.
 69. Kawatus, K.; Shibata, T.; Hamano, Y. Application of immunoaffinity chromatography for detection of tetrodotoxin from urine samples of poisoned patients. *Toxicol.* **1999**, *37*, 325–333.
 70. Pohl, H. A. *Dielectrophoresis: The Behavior of Neutral Matter in Non-uniform Electric Fields*. Cambridge University Press: Cambridge, NY; 1978.
 71. Asbury, C. L.; van den Engh, G. Trapping of DNA in nonuniform oscillating electric fields. *Biophys. J.* **1998**, *74*, 1024–1030.
 72. Ramos, A.; Morgan, H.; Green, N. G.; Castellanos, A. AC electrokinetics: a review of forces in microelectrode structures. *J. Phys. D Appl. Phys.* **1999**, *31*, 2338–2353.
 73. Hughes, M. P. AC electrokinetics: applications for nanotechnology. *Nanotechnology* **2000**, *11*, 124–132.
 74. Wong, P. K.; Wang, T.-H.; Deval, J. H.; Ho, C.-M. Electrokinetics in micro devices for biotechnology applications. *IEEE-ASME Trans. Mechatronics* **2004**, *9*, 366–376.
 75. Jones, T. B. *Electromechanics of Particles*. Cambridge University Press: Cambridge, NY; 1995.
 76. Gimsa, J. A comprehensive approach to electro-orientation, electrodeformation, dielectrophoresis, and electrorotation of ellipsoidal particles and biological cells. *Bioelectrochemistry* **2001**, *54*, 23–31.
 77. Chou, C. F.; et al. Electrodeless dielectrophoresis of single- and double-stranded DNA. *Biophys. J.* **2002**, *83*, 2170–2179.

78. Huang, Y.; et al. Dielectrophoretic cell separation and gene expression profiling on microelectronic chip arrays. *Anal. Chem.* **2002**, *74*, 3362–3371.
79. Chiou, P. Y.; Ohta, A. T.; Wu, M. C. Massively parallel manipulation of single cells and microparticles using optical images. *Nature* **2005**, *436*, 370–372.
80. Green, N. G.; Ramos, A.; Gonzalez, A.; Morgan, H.; Castellanos, A. Fluid flow induced by nonuniform AC electric fields in electrolytes on microelectrodes. I. Experimental measurements. *Phys. Rev. E Stat. Phys. Plasmas Fluids Relat. Interdiscip. Topics* **2000**, *61*, 4011–4018.
81. Gonzalez, A.; Ramos, A.; Green, N. G.; Castellanos, A.; Morgan, H. Fluid flow induced by nonuniform AC electric fields in electrolytes on microelectrodes. II. A linear double-layer analysis. *Phys. Rev. E* **2000**, *61*, 4019–4028.
82. Green, N. G.; Ramos, A.; Gonzalez, A.; Morgan, H.; Castellanos, A. A fluid flow induced by nonuniform AC electric fields in electrolytes on microelectrodes. III. Observation of streamlines and numerical simulation. *Phys. Rev. E* **2002**, *66*, 305–316.
83. Wong, P. K.; Chen, C. Y.; Wang, T. H.; Ho, C. M. Electrokinetic bioprocessor for concentrating cells and molecules. *Anal. Chem.* **2004**, *76*, 6908–6914.
84. Sin, M. L. Y.; Constantino, V. U.; Gau, V.; Haake, D. A.; Wong, P. K. Scaling analysis of a universal electrode for molecular biosensors. *IEEE Rev. Adv. Micro Nano Mol. Syst.* (accepted 2007).
85. Sin, M. L. Y.; Gau, V.; Liao, J. C.; Haake, D. A.; Wong, P. K. Active manipulation of quantum dots using AC electrokinetics. *J. Phys. Chem. C* **2009**, *113*, 6561–6565.
86. Souverain, S.; Rudaz, S.; Veuthey, J.-L. Matrix effect in LC-ESI-MS and LC-APCI-MS with off-line and online extraction procedures. *J. Chromatogr. A* **2004**, *1058*, 61–66.
87. Mei, H.; Hsieh, Y.; Nardo, C.; Xu, X.; Wang, S.; Ng, K.; Korfmacher, W. A. Investigation of matrix effects in bioanalytical high-performance liquid chromatography/tandem mass spectrometric assays: application to drug discovery. *Rapid Commun. Mass Spectrom.* **2009**, *17*, 97–103.
88. Taylor, A. W.; Bruno, R. S.; Frei, B.; Traber, M. G. Benefits of prolonged gradient separation for high-performance liquid chromatography–tandem mass spectrometry quantitation of plasma. *Anal. Biochem.* **2006**, *350*, 41–51.
89. Andrews, C. L.; et al. Improved liquid chromatography–mass spectrometry performance in quantitative analysis using a nanosplitter interface. *J. Chromatogr. A* **2004**, *1053*, 1–2.
90. Andrews, C. L.; Feng, L.; Yang, E.; Yu, C.-P.; Vouros, P. Incorporation of a nanosplitter interface into an LC-MS-RD system to facilitate drug metabolism studies. *J. Mass Spectrom.* **2006**, *41*, 43–49.
91. Juraschek, R.; Dülcks, T.; Karas, M. Nanoelectrospray—more than just a minimized-flow electrospray ionization source. *J. Am. Soc. Mass Spectrom.* **1999**, *10*, 300–308.
92. Hong, M. In *Using Mass Spectrometry for Drug Metabolism Studies*; Korfmacher, W. A., Ed.; CRC Press: Boca Raton, FL; 2005; pp 103–150.
93. Wang, S.; Cyronak, M.; Yang, E. Does a stable isotopically labeled internal standard always correct analyte response? A matrix effect study on a LC/MS/MS method for the determination of carvedilol enantiomers in human plasma. *J. Pharm. Biomed. Anal.* **2007**, *43*, 701–707.
94. Sojo, L. E.; Lum, G.; Chee, P. Internal standard signal suppression by co-eluting analyte in isotope dilution LC-ESI-MS. *Analyst* **2003**, *128*, 51–54.
95. Liang, H. R.; Foltz, R. L.; Meng, M.; Bennett, P. Ionization enhancement in atmospheric pressure chemical ionization and suppression in electrospray ionization between target drugs and stable-isotope-labeled internal standards in quantitative liquid chromatography/tandem mass spectrometry. *Rapid Commun. Mass Spectrom.* **2003**, *17*, 2815–2821.
96. Matuszewski, B. K.; Constanzer, M. L.; Chavez-Eng, C. M. Matrix effect in quantitative LC/MS/MS analyses of biological fluids: a method for determination of finasteride in human plasma at pictogram per milliliter concentrations. *Anal. Chem.* **1998**, *70*, 882–889.
97. Hsieh, Y. Potential of HILIC-MS in quantitative bioanalysis of drugs and drug metabolites. *J. Sep. Sci.* **2008**, *31*, 1481–1491.
98. Situ, C.; Wylie, A. R. G.; Douglas, A.; Elliott, C. T. Reduction of severe bovine serum associated matrix effects on carboxymethylated dextran coated biosensor surfaces. *Talanta* **2008**, *76*(4, 15), 832–836.
99. Ratcliffe, W. A.; Corrie, J. E.; Dalziel, A. H.; Macpherson, J. S. Direct 125I-Radioligand assays for serum progesterone compared with assays involving extraction of serum. *Clin. Chem.* **1982**, *28*, 1314–1318.
100. Basu, A.; Shrivastav, T. G.; Maitra, S. K. A direct antigen heterologous enzyme immunoassay for measuring progesterone in serum without using displacer. *Steroids* **2006**, *71*, 222–230.
101. Wild, D. In *Immunoassay Handbook*; Wild, D., Ed.; 3rd ed. Elsevier: Oxford, UK; 2005; pp 141.
102. Ullman, E. Homogeneous immunoassays: historical perspective and future promise. *J. Chem. Ed.* **1999**, *76*, 781–788.
103. Marchi, I.; Rudaz, S.; Veuthey, J.-L. Sample preparation development and matrix effects evaluation for multianalyte determination in urine. *J. Pharm. Biomed. Anal.* **2009**, *49*(2, 20), 459–467.
104. Möller, K.; Nilsson, U.; Crescenzi, C. Investigation of matrix effects of urine on a molecularly imprinted solid-phase extraction. *J. Chromatogr. B* **2004**, *811*(2, 25), 171–176.
105. Mach, K. E.; Du, C. B.; Phull, H.; Haake, D. A.; Shih, M. C.; Baron, E. J.; Liao, J. C. Multiplex pathogen identification for polymicrobial urinary tract infections using biosensor technology: a prospective clinical study. *J. Urol.* **2009**, *182*(6), 2735–2741.
106. Damsa, R.; Huestis, M. A.; Lambert, W. E.; Murphy, C. M. Matrix effect in bio-analysis of illicit drugs with LC-MS/MS: influence of ionization type, sample preparation, and biofluid. *J. Am. Soc. Mass Spectrom.* **2003**, *14*(11), 1290–1294.
107. Saar, E.; Gerostamoulos, D.; Drummer, O. H.; Beyer, J. Comparison of extraction efficiencies and LC–MS–MS, matrix effects using LLE and SPE methods for 19 antipsychotics in human blood. *Anal Bioanal Chem* **2009**, *393*, 727–734.
108. Mitchell, J. S.; Lowe, T. E. Matrix effects on an antigen immobilized format for competitive enzyme immunoassay of salivary testosterone. *J. Immunol. Methods* **2009**, *349*(1–2, 30), 61–66.
109. Mitchell, J. S.; Lowe, T. E.; Ingram, J. R. Rapid ultrasensitive measurement of salivary cortisol using nano-linker chemistry coupled with surface plasmon resonance detection. *Analyst* **2009**, *134*(2), 380–386 Epub November 13, 2008.
110. Gaster, R. S.; Hall, D. A.; Nielsen, C. H.; Osterfeld, S. J.; Yu, H.; Mach, K. E.; Wilson, R. J.; Murmann, B.; Liao, J. C.; Gambhir, S. S.; Wang, S. X. Matrix-insensitive protein assays push the limits of biosensors in medicine. *Nat. Med* **2009**, *15*(11), 1327–1332.